

Cu(II)-All Solid State Sensor Ion Selective Electrode (ASS-ISE) with a Nano-molar Detection Limit and its Use for the Analysis of Waste Water Samples

M. R. Ganjali^{1,2,*}, M. Rezapour³, M. Pirali-Hamedani^{4,5}, H. Rashedi⁶

¹Center of Excellence in Electrochemistry, Faculty of Chemistry, University of Tehran, Tehran, Iran

²Biosensor Research Center, Endocrinology & Metabolism Molecular-Cellular Sciences Institute, Tehran University of Medical Sciences, Tehran, Iran

³IP Department, Research Institute of Petroleum Industry (RIPI), P.O. Box 14665-137, Tehran, Iran

⁴Department of Medicinal Chemistry, Faculty of Pharmacy, Tehran University of Medical Sciences, Tehran, Iran

⁵Pharmaceutical Sciences Research Center, Tehran, Iran

⁶Department of Chemical Engineering, Faculty of Engineering, University of Tehran, Tehran, Iran

*E-mail: ganjali@gmail.com; ganjali@khayam.ut.ac.ir

Received: 29 May 2015 / Accepted: 9 July 2015 / Published: 28 July 2015

The present work focuses on the development of an all-solid-state electrode (ASSE) for the determination of Cu²⁺ ions in complex matrices. The ASSE is made of a layer of a conductive graphite-epoxy resin, coated on a copper wire. The composite layer is further covered with a Cu²⁺ selective PVC membrane composed of 33% PVC as the base material, 58% nitrobenzene (NB) as a plasticizing solvent, 3% of potassium tetrakis (p-chlorophenyl) borate (KpCITPB) which acts as an ionic additive, and 6% of a copper selective ionophore (L). The ASSE was found to have a Nernstian potential response of 29.4±0.4 mV/decade over a dynamic linear range from 1.0×10⁻⁸ M to 1.0×10⁻³ M of Cu²⁺ and the detection limit of the sensing device was also evaluated to be 2.4×10⁻⁹ mol L⁻¹. The selectivity of the sensor towards the target ion in the presence of different commonly occurring interfering ions was found to be very good, and to test its applicability in the analysis of real samples with complex matrices the ASSE was applied for the determination of Cu²⁺ in waste water samples of copper electroplating, and the results were found to be very satisfactory.

Keywords: All solid state, Potentiometry, Copper, Sensor, Waste water

1. INTRODUCTION

Rather large concentrations of Cu²⁺ ions are present in sea water and the element has proven to influence the lifetime of phytoplankton [1,2]. Cu²⁺ also has critical roles in animal husbandry and

human health, and even trace amounts of the ion can have toxicological effects on various organisms [3, 4], the analysis of the ion is an important analytical task. The importance is further highlighted due to the increased levels of Cu^{2+} in the environment due to its emission from various industrial sources, which pronounce the dangers posed to public health [5], and in turn enhances the significance of monitoring even the trace amounts of Cu^{2+} has become more and more important because of the increased interest in environmental pollution [6].

Other reasons behind the importance of monitoring in biological and environmental systems is its role in the absorption and function of iron especially in the process of production of hemoglobin, as well as the role Cu^{2+} in enzymes with fundamental effects in preventing the oxidation, and manufacturing collagen and generating energy [7].

It is clear that sensitive, reproducible and accurate analytical methods are always the basic need for the evaluation of any analyte, Cu^{2+} included, in different samples [8]. The conventional methods used to this end, in the case of copper ions include instrumental techniques like spectrophotometry [9,10] dispersive liquid–liquid microextraction [11,12], adsorptive stripping voltammetry [13], sequential injection analysis [14], high performance liquid chromatography [15], anodic stripping voltammetry [16, 17], but the advantages of ion selective electrodes (ISEs) including accuracy, portability, sensitivity and selectivity, low cost and ease of use have led to the development and application of some ISEs based on the application of different ionophores, for the determination of Cu^{2+} ions [18-24].


Advantages of ASS-ISEs [25], including absence of limitations associated with the presence of an internal solution in symmetrical devices leading to limitations in the limit of detection, life time and applications position of the devices, in addition to the possibility of using different intermediary materials in the construction of the devices have attracted an increasing interest to these devices, in recent years [26-37]. Other more distinct advantages of all solid-state ion-selective electrodes (ASS-ISEs), include their modified detection limits and mechanical stability. The former, as mentioned earlier, arises from the elimination of the inner filling reference solution which in turn the need for its optimization and leaking [38–40]. The latter advantage provides the possibility of more flexible mechanical design of the electrodes as well as production of disposable ones.

Given the advantages of polymeric membrane ISEs, and the need of precise, accurate and selective determination of low amounts of Cu^{2+} in water samples, different ligands have been tested and some have been to act selective ionophores for the construction of polymeric based copper-selective electrodes [41]. There is no need to mention that devising a portable, yet powerful device with a low detection limit and easy application procedure, for the online monitoring of Cu^{2+} concentration in diverse samples, which can be achieved through the development of an ASS-ISE device can be of great interest for quality control, environmental and biological applications.

So the present work focuses on the design and construction of a novel copper selective ASS-ISE, based on the application of a conductive polymeric composite (CPC), composed of multi-walled carbon nanotubes (MWCNTs) and epoxy resin as the internal contact and transducer. The CPC layer, which is internally in contact with a copper wire, is then coated with a layer of the PVC membrane to furnish it with copper selectivity required for the construction of an ion selective electrode.

2. MATERIALS AND METHODS

Tetrahydrofuran (THF), dibutyl phthalate (DBP), nitrobenzene (NB), benzyl acetate (BA), *o*-nitrophenyloctylether (NPOE), oleic acid (OA), potassium tetrakis (*p*-chlorophenyl) borate (KpCITPB) and sodium tetrapenyl borate (NaTPB), graphite powder (1–2 μm particle size), the MWCNTs and all salts (nitrate or chloride form) were of analytical reagent grade and were obtained from Merck Co., High-molecular weight polyvinylchloride (PVC) was from Fluka Co. and the ionophore L (Scheme 1) was synthesized as described earlier [42]. Macroplast Su 2227 epoxy was from Henkel Co. and the desmodur RFE hardener was obtained from Bayer Ag.



Scheme 1. Chemical structure of the used selectophore

To prepare the CPC which is a MWCNTs-loaded epoxy resin in nature, and is used as the internal contact and transducer, different proportions of graphite powder, MWCNTs, epoxy, and hardener were mixed in a THF solvent, let to rest for about 20-30 minutes in air and tested. After the mentioned period which was marked by reaching a viscose mixture, a section of a shielded copper wire (0.5 mm diameter and 15 cm length) was polished and soaked into the viscose solution so as to be coated with it and was left to dry for half a day.

The solid-state contact prepared was next immersed in the PVC membrane cocktail three or four times to be covered with the polymeric membrane, and then left in the air for 24 hours to dry. The optimal PVC membrane was experimentally found to compose 6% w/w of L, 58% w/w of the optimal plasticizer (NB), 33% w/w of PVC, and 3% w/w of KpCITPB. The PVC membrane cocktails were by dissolving various amounts of the ligand L with appropriate amounts of PVC, plasticizer and ionic additive in THF, thoroughly mixing the ingredients transferring the resulting mixture to a glass dish of 2 cm diameter. The cocktails were next left to viscosity in air, and the coated on the solid-state contact as mentioned above. The ASS-ISEs prepared according to the above-described procedure were finally conditioned in a 10^{-3} mol L⁻¹ solution of Cu²⁺.

The ASS-ISEs were next used as an indicator electrode in a potentiometric cell consisting of an Ag/AgCl double junction reference electrodes (Azar-Electrode Co., Iran) as the external reference electrode. The potentiometric cell can hence be illustrated as bellow:

Cu wire/ASS layer/ion selective PVC membrane | sample solution || Ag-AgCl, KCl (satd.)

An ion analyzer with a 250 pH/mV meter with ± 0.1 mV precision was used to connect the two electrodes in all experiments, and the measurements were carried out through the calibration method.

3. RESULTS AND DISCUSSION

Interestingly, the majority of previously reported ISEs are symmetrical membranes where the polymeric ion selective electrode interfaces the internal standard and external test solutions. In asymmetrical ISEs like ASS-ISE devices [43-46], however, only the exterior side of the ISE membrane contacts with the test solution while the internal surface is in adheres to the solid-state contact.

Table 1. Various membrane ingredients used in making ASS-PVC membrane sensor

No	Composition (%)				Slope* (mV/decade)	DL (mol/L)
	PVC	Plasticizer	Ionophore	Additive		
1	33	63 DBP	4	-	16.2 \pm 0.5	3.7 \times 10 ⁻⁸
2	33	61 DBP	6	-	17.4 \pm 0.4	3.5 \times 10 ⁻⁸
3	33	60 DBP	7	-	17.3 \pm 0.4	3.5 \times 10 ⁻⁸
4	33	59 DBP	6	2 KpCITPB	25.4 \pm 0.6	8.5 \times 10 ⁻⁹
5	33	58 DBP	6	3 KpCITPB	25.8 \pm 0.3	8.5 \times 10 ⁻⁹
6	33	51 DBP	6	10 OA	26.1 \pm 0.5	1.4 \times 10 ⁻⁸
7	33	46 DBP	6	15 OA	26.0 \pm 0.3	1.5 \times 10 ⁻⁸
8	33	41 DBP	6	20 OA	26.0 \pm 0.4	7.1 \times 10 ⁻⁹
9	33	58 BA	6	3 KpCITPB	26.7 \pm 0.3	5.5 \times 10 ⁻⁹
10	33	58 NB	6	3 KpCITPB	29.4 \pm 0.4	2.4 \times 10 ⁻⁹
11	33	58 NPOE	6	3 KpCITPB	29.8 \pm 0.5	3.2 \times 10 ⁻⁹
12	33	74 NPOE	-	3 KpCITPB	5.4 \pm 0.6	3.5 \times 10 ⁻⁷

*The results are based on five replicate measurements.

The inner solution in symmetrical devices helps develop the potential gradient across the membrane, yet it causes limitations mentioned above especially in terms of robustness and miniaturization, which are greatly improved by the advent of asymmetrical devices providing them

with the capabilities required for the applications in medicinal, biological, environmental and even in vivo systems. The application of conductive epoxy supports in ASS-ISEs allows for the use of PVC membrane, with variable and adjustable compositions, which acts as the main components of the sensing device. There is no need to mention that each component of the PVC membrane and its absolute and relative amounts in the overall composition plays a distinct part in the response of the resulting sensor [47-53].

As a result the important information on the optimization of the membrane composition is given in Table 1, to provide the reader with better insight. It has been established in the literature that generally a plasticizer/PVC ratio of around 2.0 leads to optimal responses by PVC membranes, and consequently, for all compositions illustrated in Table 1 the ratio is true. Hence the amount of the PVC, i.e. 33%wt is not mentioned in any case. The plasticizer, is used to enhance the mobility of the free and complexed ligand throughout the membrane.

The materials used as the plasticizer need to be very non- volatile and able to physically dissolve PVC, possess no chemically active functional groups and have proper polarity (which is determined by the nature of the target species). Conventional plasticizers of various dielectric constants including dibutyl phthalate (DBP), nitrobenzene (NB), *o*-nitrophenyloctylether (NPOE) and benzylacetate (BA) were tested in this work and NPOE and NB led to better responses, which can be attributed to the relatively high polarity of Cu^{2+} ions which prefers a high polarity medium.

A third generally used, yet very critical membrane ingredient is the ionic additive. Although the proper polarity of the plasticizer facilitates the extraction and exchange of the target ions into the membrane, incorporation of small amounts of such ionic additives has been found not only to improve the exchange mechanism but also to reduce the Ohmic resistance of the PVC membrane.

The effect of the amount of the ionophore on the potential response of the sensor was depicted in Table 1. As it can be seen from Table 1, 6% L is the best amount of ionophore (membrane no. 2). The response of the sensor was also evaluated in absence of the ionophore. As can be seen, a blank membrane, the membrane without the sensing ionophore, responds too poor (membrane no. 12). Hence, membrane no. 10 showed the best Nernstian slope (29.4 ± 0.4 mV/decade) was selected for further studies.

Further evaluation of the ASS-ISEs included studying their potential response at different Cu^{2+} concentration (i.e. from 1.0×10^{-9} to 1.0×10^{-1} mol L^{-1}) and drawing the potential vs. $-\log [\text{Cu}^{2+}]$. The response attributed to the sensor with the optimal PVC membrane concentration is illustrated in Figure 1, the linear segment of which (i.e. from 1.0×10^{-8} - 1.0×10^{-3} mol L^{-1}) represents the linear range of the sensor.

Although the conventional PVC membrane sensors have linear ranges in the range of 0.1 to 10^{-5} or even 10^{-6} mol L^{-1} , the asymmetrical ASS-ISE membrane sensor led to far lower detection limits in the order of sub 10^{-8} mol L^{-1} (i.e. 2.4×10^{-9} mol L^{-1} , achieved through the extrapolation of the two linear segments of the lower part of the calibration curves). The upper detection limit of the electrode was observed to be 10^{-3} mol L^{-1} , which was attributed to the saturation of its active at this concentration, which is higher than those of the symmetrical sensors that have a larger surface area and hence have improved upper detection limits.




Figure 1. Calibration curve of Cu ASS PVC membrane sensor; the results are based on 5 replicate measurements.

To further evaluate the ASS-ISE its dynamic response time, defined as the time required for the response to reach ± 1 mV of its equilibrium potential upon a 10 fold increase in the concentration of the test solution [54-55] created by its successive immersion into various sample solutions in the range of 1.0×10^{-8} to 1.0×10^{-3} mol L⁻¹. The results (Figure 2) revealed the sensor to be able to quickly reach the equilibrium response in the whole range and the average time for doing, or the response time of the ASS-ISE membrane sensor was evaluated to be about 7 s.

The effect of sample solution pH on the response is also crucial in evaluating a potentiometric sensor. To study the range in which the response of the ASS-ISE is independent of pH, the response of the sensor towards a fixed and known concentration of Cu²⁺ (1.0×10^{-5} mol L⁻¹) was monitored while altering the pH of the sample solution from 2.0 to 10.0 (the alterations were done using concentrated NaOH or HCl solutions to avoid considerable volume and concentration changes). The results (Fig. 3) showed that no considerable potential changes occur upon changing the pH in the range of 4.0 to 8.5. The rather considerable potential fluctuations on the two sides of the range render the device not applicable outside it. Above 8.5 pH values lead to the formation of Cu(OH)⁺ or Cu(OH)₂, lowering the concentration of the free Cu²⁺, while below 4.0, the rather acidic media causes the protonation of the nitrogen atoms of L.




Figure 2. Dynamic response time of the Cu ASS PVC membrane sensor




Figure 3. pH effect on potential response of the sensor

The lifetime of the sensor was also evaluated by recording its calibration curve and potential response, through daily tests conducted using standard solutions. During each test three ASS-ISEs were used for 1 hour (for 12 weeks). The results showed that the average lifetime of the optimal ASS-ISEs falls within 4–10 weeks [56-59]. After 10 weeks of use under he mentioned

conditions, a gradual decrease in the slope and some increase in the LOD were detected. This can be attributed to the leakage and loss of the plasticizer, the ionophore, or the ionic additive from the PVC membrane due to extensive use.

The most important characteristic of the ASS-ISEs (like any other ion selective electrode), i.e. their selectivity was also evaluated. This parameter is defined as the ISEs' specific response to the target ion in the presence of interfering ions. The method applied to the evaluation of this parameter was the so-called matched potential method (MPM) [60-63] and the resulting selectivity coefficients are given in Table 2. The results show that none of the tested interfering ions has a considerable influence on the response of the sensor.

Table 2. The selectivity coefficients of various interfering cations for Cu^{2+} ASS PVC membrane sensor

Cation	Selectivity Coefficients
Na^+	$<10^{-6}$
K^+	$<10^{-6}$
Mg^{2+}	$<10^{-6}$
Ca^{2+}	$<10^{-6}$
Hg^{2+}	5.2×10^{-4}
Zn^{2+}	2.0×10^{-4}
Co^{2+}	3.3×10^{-4}
Cd^{2+}	7.3×10^{-4}
Pb^{2+}	6.8×10^{-4}
Ag^+	2.7×10^{-4}
Fe^{3+}	3.9×10^{-4}
Ni^{2+}	4.1×10^{-4}
La^{3+}	3.0×10^{-4}
Al^{3+}	2.1×10^{-4}
Ce^{3+}	2.6×10^{-4}

To prove that the ASS-ISE can be applied to real samples, the concentration of Cu^{2+} in some industrial wastewater samples was evaluated using the device. The samples were only acidified with HNO_3 after acquisition, and three analyses were performed on each sample, through the calibration method. The results were validated using the results of the analyses of the same samples by inductively coupled plasma optical emission spectrometry (ICP-OES). The test results (Table 3), show the amount of copper with the help of the sensor and are in good agreement with reference method.

Table 3. Results of copper analysis in waste water samples

Sample	Found by the sensor* (ppm)	Found by ICP-OES (ppm)
Sample 1	1.7±0.1	1.6±0.1
Sample 2	3.4±0.2	3.3±0.1
Sample 3	0.9±0.1	0.9±0.1

* The results are based on five replicate measurements.

4. CONCLUSION

An All-solid-state ion selective electrode (ASS-ISE) potentiometric PVC membrane sensor has been devised and used for facile and rapid determination of Cu^{2+} in waste water samples. The solid conductive polymeric composite (CPC) used as the transducer, composed of graphite powder, MWCNTs, and epoxy resin and was coated on a copper wire. The overall solid contact was then covered with a thin layer PVC membrane, composed of 33% PVC, 58% NB, 3% ionic additive, and 6% ionophore L. The ASS-ISE had a dynamic linear range from 1.0×10^{-8} to 1.0×10^{-3} M of Cu^{2+} and the detection limit was evaluated at 2.4×10^{-9} M. The proposed ASS-ISE was successfully applied to the analysis of Cu^{2+} in waste water samples.

ACKNOWLEDGEMENTS

Authors are grateful to the research council of University of Tehran for the financial support of this work.

Reference

1. A. Zirino, D.A. VanderWeele, S.L. Belli, R. Demarco, and D.J. Mackey, *Mar. Chem.*, 61 (1998) 173.
2. R.S. Eriksen, D.J. Mackey, R. van Dam, and B. Nowak, *Mar. Chem.*, 74 (2001) 99.
3. I. Durukan, C.A. Şahin, and S. Bektaş, *Microchem. J.* 98 (2011) 215
4. J. Schwarcz, and J.A. Schwarcz, *The Fly in the Ointment: 70 Fascinating Commentaries on the Science of Everyday Life*; ECW Press: Toronto, Canada (2004) p. 180.
5. A.M. Jamaluddin, M. Saifuddin, T. Jannat, and S.C. Bhattacharjee, *Green Page* (2010), available online at: <http://www.eco-web.com/edi/100412.html>.
6. M.R. Shishehbore, N. Nasirizadeh, S.A.M. Haji, and M. Tabatabaee, *Can. J. Anal. Sci. Spectrosc.* 50 (2005) 130.
7. [online] available at: <http://www.nutrition.org/nutinfo/content/copp.shtml>.
8. S. Jignesh, K. Vineeta, S. Abhay, and K. Vilasrao, *Int. J. Res. Pharm. Chem.* 2 (2012) 146.
9. P.H. Ahmad, M. Karimi, E. Moniri, and H. Soudi, *Afr. J. Pure Appl. Chem.* 2 (2008) 096.
10. J. Karthikeyan, P.P. Naik, and A.N. Shetty, *Environ. Monit. Assess.* 176 (2010) 419.
11. J. Škrliková, V. Andruch, I.S. Balogh, L. Kocúrová, L. Nagy, and Y. Baze, *Microchem. J.* 99 (2011) 40.
12. K. Shrivastava, *Environ. Monit. Assess.* 168 (2010) 315.

13. S. Abbasi, H. Khani, and R. Tabaraki, *Food Chem.* 123 (2010) 507.
14. T. Leelasattarakul, S. Liawruangrath, M. Rayanakorn, W. Oungpipat, and B. Liawruangrath, *Talanta* 70 (2006) 656.
15. V. Kaur, and A.K. Malik, *Ann. Chim.* 97 (2007) 1279.
16. T. Rohani, and M.A. Taher, *J. AOAC Int.* 91 (2008) 1478.
17. B.C. Janegitz, L.H. Marcolino-Junior, S.P. Campana-Filho, R.C. Faria, and O. Fatibello-Filho, *Sens. Actuators B* 142 (2009) 260.
18. V.K. Gupta, A.K. Singh, M. Al Khayat, Barkha Gupta, *Anal. Chim. Acta*, 590 (2007) 81.
19. H. A. Zamani, M. Mohammadhosseini, S. Haji-Mohammadrezazadeh, F. Faridbod, M. R. Ganjali, S. Meghdadi and A. Davoodnia, *Mater. Sci. Eng. C*, 32 (2012) 712.
20. V. K. Gupta, A. K. Singh and B. Gupta, *Anal. Chim. Acta*, 575 (2006) 198.
21. M. R. Ganjali, P. Norouzi, A. Atrian, F. Faridbod, S. Meghdadi, and M. Giahi, *Mater. Sci. Eng. C*, 29 (2009) 205.
22. H. A. Zamani, M. R. Ganjali, P. Norouzi, and M. Adib, *Mater. Sci. Eng. C*, 28 (2008) 157.
23. M. R. Ganjali, A. Karimi, S. J. Shahtaheri, and P. Norouzi, *Int. J. Electrochem. Sci.*, 8 (2013) 1999.
24. A. K. Singh, V. K. Gupta and B. Gupta, *Anal. Chim. Acta*, 1 (2007) 171.
25. B.P. Nikolskii, and E.A. Materowa, *Ion. Sel. Electrode Rev.*, 7 (1985) 3.
26. J.M.C.S. Magalhaes, F. Cespedes, S. Alegret, and A.A.S.C. Machado, *Anal. Chim. Acta* 355 (1997) 241.
27. H.J. Yoon, J.H. Shin, S.D. Lee, H. Nam, G.S. Cha, T.D. Strong, and R.B. Brown, *Sens. Actuators B* 64 (2000) 8.
28. S. Walsh, D. Diamond, J. McLaughlin, E. McAdams, D. Woolfson, D. Jones, and M. Bonner, *Electroanalysis* 9 (1997) 1318
29. O. Lutze, R.K. Meruva, A. Frielich, N. Ramamurthy, R.B. Brown, R. Hower, and M.E. Meyerhoff, *Fresenius J. Anal. Chem.* 364 (1999) 364
30. A. Cadogan, Z. Gao, A. Lewenstam, A. Ivaska, and D. Diamond, *Anal. Chem.* 64 (1992) 2496
31. A. Michalska, A. Hulanicki, and A. Lewenstam, *Microchem. J.* 57 (1997) 59
32. T. Lindfors, and A. Ivaska, *Anal. Chim. Acta* 400 (1999) 101
33. T. Lindfors, and A. Ivaska, *Anal. Chim. Acta* 404 (2000) 111
34. J. Bobacka, *Anal. Chem.* 71 (1999) 4932
35. P. Sjoberg, J. Bobacka, A. Lewenstam, and A. Ivaska, *Electroanalysis* 11 (1999) 821
36. F. Song, J.H. Ha, B.H. Park, T.H. Kwak, I.T. Kim, H.N. Nam, and G.S. Cha, *Talanta* 57 (2002) 263
37. N. Tinkilic, O. Cubuk, and I. Isildak, *Anal. Chim. Acta* 452 (2002) 29.
38. M. R. Ganjali, Z. Memari, F. Faridbod, and P. Norouzi, *Int. J. Electrochem. Sci.* 3 (2008) 1169.
39. C.Z. Lai, M.M. Joyer, M.A. Fierke, N.D. Petkovich, A. Stein, and P. Bühlmann, *J. Solid State Electrochem.* 13 (2009) 123.
40. K. Wang, J.J. Xu, K.S. Tang, and H.Y. Chen, *Talanta* 67 (2005) 798.
41. S. Sadeghi, and F. Fathi, *J. Incl. Phenom. Macrocycl. Chem.* 67 (2010) 91.
42. M. R. Ganjali, M. Emami and M. Salavati-Niasari, *Bull. Korean Chem. Soc.* 23 (2002) 1394.
43. M.R. Ganjali, F. Faridbod, N. Davarkhah, S.J. Shahtaheri, and P. Norouzi, *Int. J. Environ. Res.* 9 (2015) 333.
44. F. Faridbod, M. R. Ganjali, B. Larijani, P. Norouzi, S. Riahi and F. S. Mirnaghi, *Sensors* 7 (2007) 3119.
45. M. R. Ganjali, P. Norouzi, F. S. Mirnaghi, S. Riahi, and F. Faridbod, *IEEE Sensors J.* 7 (2007) 1138.
46. M. R. Ganjali, N. Motakef-Kazami, F. Faridbod, S. Khoei, and P. Norouzi, *J. Hazard. Mater.* 173 (2010) 415.
47. F. Faridbod, F. Mizani, M. R. Ganjali, and P. Norouzi, *Int. J. Electrochem. Sci.* 8 (2013) 10461.
48. M. R. Ganjali, B. Larijani, F. Faridbod and P. Norouzi, *Int. J. Electrochem. Sci.* 8 (2013) 10487.

49. F. Faridbod, M. R. Ganjali and P. Norouzi, *Int. J. Electrochem. Sci.* 8 (2013) 6107.
50. H.A. Zamani, M. Nekoei, M. Mohammadhosseini, and M.R. Ganjali, *Mater. Sci. Eng. C*, 30 (2010) 480.
51. M. Javanbakht, S. E. Fard, A. Mohammadi, M. Abdouss, M. R. Ganjali, P. Norouzi, and L. Safaraliev, *Anal. Chim. Acta*, 612 (2008) 65.
52. M. R. Ganjali, M. Qomi, A. Daftari, P. Norouzi, M. Salavati-Niasari and M. Rabbani, *Sens. Actuators B* 98 (2004) 92.
53. H. A. Zamani, G. Rajabzadeh and M. R. Ganjali, *Sensor Lett.* 7 (2009) 114.
54. H. A. Zamani, M. R. Ganjali, P. Norouzi and M. Adib, *Sensor Lett.* 5 (2007) 522.
55. M. Shamsipur, S. Rouhani, H. Shaghi, M. R. Ganjali, and H. Eshghi, *Anal. Chem.* 71 (1999) 4938.
56. M. Javanbakht, M. R. Ganjali, P. Norouzi, A. Badiei, A. Hasheminasab, and M. Abdouss, *Electroanalysis* 19 (2007) 1307.
57. H. A. Zamani, M. R. Ganjali, and M. Adib, *Sensor Lett.* 6 (2006) 345.
58. M. R. Ganjali, P. Norouzi, F. Faridbod, S. Riahi, J. Ravanshad, J. Tashkhourian, M. Salavati-Niasari and M. Javaheri, *IEEE Sensors J.* 7 (2007) 544.
59. M. R. Ganjali, P. Norouzi, M. Adib and A. Ahmadalinezhad, *Anal. Lett.* 39 (2006) 1075.
60. H. A. Zamani, J. Abedini-Torghabeh, and M. R. Ganjali, *Electroanalysis* 18 (2006) 888.
61. H. A. Zamani, M. R. Ganjali, P. Norouzi, and S. Meghdadi, *Anal. Lett.* 41 (2008) 902.
62. H. A. Zamani, G. Rajabzadeh, M. Masrornia, A. Dejbord, M. R. Ganjali, and N. Seifi, *Desalination*, 249 (2009) 560.
63. M. Hosseini, S. D. Abkenar, M. R. Ganjali and F. Faridbod, *Mater. Sci. Eng. C*, 31 (2011) 428.

© 2015 The Authors. Published by ESG (www.electrochemsci.org). This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution license (<http://creativecommons.org/licenses/by/4.0/>).